

## CERTIFICATE OF ANALYSIS

### MRX D-Dimer Reagent

Product number: MRX147B OEM

Lot number: 19763

Expiry Date: 2020 08

<b>Analysis</b>	<b>Specification</b>	<b>Results</b>
Reagent sensitivity, slope	0,155-0,189	0,177
Precision ( $r^2$ )	$\geq 0,99$	0,998

This is to certify that the above product is performing according to the Quality Control specifications of MediRox.

Sofia Wennerberg Eriksson, QC, MediRox 2019-03-26