

CERTIFICATE OF ANALYSIS

MRX D-Dimer Reagent

Product number: MRX147B OEM

Lot number: 19759

Expiry Date: 2020-08-25

Analysis	Specification	Results
Reagent sensitivity, slope	0,155-0,189	0,177
Precision (r^2)	$\geq 0,99$	0,998

This is to certify that the above product is performing according to the Quality Control specifications of MediRox.

Linda Carlsson, QC, MediRox 2019-03-20